

SPECIFICATIONS

NI VB-8012

NI VirtualBench™ All-In-One Instrument

Français	Deutsch	日本語	한국어	简体中文
ni.com/manuals				

These specifications are for the National Instruments VirtualBench VB-8012 only. These specifications are valid following 30 minutes of warmup and are typical at 25 °C unless otherwise noted.

Mixed Signal Oscilloscope

Analog Channels

Vertical System

Number of channels.....	2 single-ended, non-isolated
Bandwidth (-3 dB) ¹	100 MHz
Resolution.....	8 bits
Accuracy ²	±2% of input, ±1% full scale
Input coupling.....	DC, AC

¹ Bandwidth using the accessory oscilloscope probe in 10X mode.

² Indicates warranted specifications valid at $T_{cal} \pm 5$ °C. Temperature coefficients are calculated using the temperature change from last external calibration.

Vertical sensitivity (range)

- 10 mV/div (100 mV_{pk-pk})
- 20 mV/div (200 mV_{pk-pk})
- 50 mV/div (400 mV_{pk-pk})
- 100 mV/div (1 V_{pk-pk})
- 200 mV/div (2 V_{pk-pk})
- 500 mV/div (4 V_{pk-pk})
- 1 V/div (10 V_{pk-pk})
- 2 V/div (20 V_{pk-pk})
- 5 V/div (40 V_{pk-pk})

Input impedance.....1 MΩ || 20 pF

Table 1. DC Offset Range

Range	Programmable Offset Range
10 mV/div, 20 mV/div, 50 mV/div	±5 V
100 mV/div, 200 mV/div, 500 mV/div, 1 V/div, 2 V/div, 5 V/div	±20 V

Acquisition modes.....Sample, peak detect, averaging

Horizontal System

Maximum sample rate.....1 GS/s single channel,
500 MS/s/channel, dual channel

Maximum record length.....1 MS/channel

Digital Channels/Logic Analyzer

Vertical System

Number of channels.....	34
Maximum input frequency.....	100 MHz
Input voltage.....	0 to 5 V



Note Mixed signal oscilloscope digital channels are designed to withstand accidental overvoltage from signals on the VB-8012 or similar devices. They are not recommended for use with signals likely to exceed 0 to 5 V in normal operation.

Input threshold.....	Programmable, 0 to 2.0 V
Threshold accuracy.....	350 mV
Additional/internal channels.....	Digital I/O lines, function generator start, external trigger (TRIG), power line frequency

Horizontal System

Timing mode sample rate ²	1 GS/s (down to ~15 kS/s)
Maximum external sample clock rate.....	100 MHz
Record length	
Typical.....	1 MS
Minimum ³	4 kS
Decimation.....	External Sample Clock, 1:1, 2:1, and n*4:1 where n is an integer
Maximum sample compression.....	2 ¹⁵ to 1

Triggering

Trigger modes.....	Normal, Auto, Single, Force
Trigger sources.....	Oscilloscope analog channels, oscilloscope digital channels, function generator start, digital I/O lines, external trigger (TRIG), power line frequency

³ Under most conditions, the logic analyzer can acquire 1 MS of data. Under some conditions with very high sustained activity on multiple inputs, the logic analyzer may only capture 4 kS of data.

Trigger types

Analog.....	Edge with hysteresis
Digital.....	Edge, glitch ⁴ , level, pattern

Trigger resolution

1 GS/s mode.....	1 ns
500 MS/s mode.....	2 ns
Trigger export.....	Available through external trigger (TRIG)

Waveform Measurements

Oscilloscope time ⁵	Period, frequency, positive duty cycle, negative duty cycle, positive pulse width, negative pulse width, rise time, fall time, rise rate, fall rate
Oscilloscope voltage ⁵	High, low, amplitude, maximum, minimum, peak-to-peak, overshoot, undershoot, RMS, mean, cycle RMS, cycle mean
Logic analyzer time ⁵	Period, frequency, positive duty cycle, negative duty cycle, positive pulse width, negative pulse width

Waveform Math

Operations ⁶	A + B, A - B, A * B, FFT
-------------------------------	--------------------------

Function Generator (FGEN)

Waveforms.....	Sine, square, ramp/triangle, DC, arbitrary ⁷
Update rate.....	125 MS/s
Resolution.....	14 bits
Number of channels.....	1
Output impedance.....	50 Ω

⁴ Glitch triggers are only available with NI VirtualBench application development support.

⁵ Waveform measurements are only available in the VirtualBench application.

⁶ Waveform math is only available in the VirtualBench application.

⁷ Arbitrary waveforms are only available with NI VirtualBench application development support.

Sine

Maximum frequency.....	20 MHz
Total harmonic distortion	
1 MHz.....	-55 dBc
10 MHz.....	-50 dBc
Spurious free dynamic range.....	-70 dB at 1 MHz (non-harmonic)
Phase noise (1 MHz).....	-115 dBc/Hz at 10 kHz offset

Square

Maximum frequency.....	5 MHz
Rise/fall time.....	<20 ns (10% to 90%)
Overshoot.....	<5%
Jitter.....	8 ns cycle-to-cycle

Ramp/triangle maximum frequency..... 150 kHz

Accuracy (with >10 k Ω load)

Amplitude (1 kHz sine).....	$\pm(1\%$ of setting + 5 mV)
DC.....	$\pm(1\%$ of setting + 5 mV)

Output range

50 Ω	± 6 V
Hi-Z (>10 k Ω).....	± 12 V

DC offset

50 Ω	± 6 V
Hi-Z (>10 k Ω).....	± 12 V



Note The combination of signal amplitude and DC offset cannot exceed the output range specifications.

Frequency

Accuracy.....	≤ 100 ppm
Resolution.....	1 μ Hz

Arbitrary waveform

Points.....	1 MS
Sample rate.....	125 MS/s

Flatness..... ± 0.3 dB to 20 MHz

Protection..... Short-circuit protected

Triggering

Trigger types.....	Start of buffer ⁸
Trigger resolution.....	8 ns
Trigger export.....	Available through external trigger (TRIG)

Digital I/O

Number of channels.....	8
Direction control.....	Input or output, software-selectable
Logic level.....	5 V compatible LVTTTL input, 3.3 V LVTTTL output
Drive strength.....	4 mA
Input voltage.....	0 to 5 V



Note Digital I/O lines are designed to withstand accidental overvoltage from signals on the VB-8012 or similar devices. They are not recommended for use with signals likely to exceed 0 to 5 V in normal operation.

Digital I/O channels 0 to 7.....	10 k Ω , pull-down pull resistors
----------------------------------	---

External Power

3.3 V output	
Voltage.....	3.3 V \pm 10%
Current.....	20 mA

Digital Multimeter

Functions.....	DC voltage, AC voltage, DC current, AC current, resistance, diode, continuity
Resolution.....	5½ digits
Sample rate.....	5 S/s

⁸ The function generator can only produce a trigger.



Caution Do not use this device for connection to signals or for measurements within Measurement Categories III or IV. For more information on Measurement Categories, refer to the *Safety Voltages* section.

Input protection

Resistance, diode.....Up to 300 V DC
 DC and AC voltage.....Up to 300 V DC or 265 V AC_{rms},
 400 V AC peak

DC and AC current

DMM A current.....Internal ceramic fuse, 10 A 250 V, time-delay,
 connector fuse 5 × 20 mm, T 10A H 250V
 (Cooper Bussmann part number S505H-10-R
 at www.cooperbussmann.com)
 DMM mA current.....Internal ceramic fuse, 1.25 A 250 V,
 connector fuse time-delay, 5 × 20 mm, T 1.25A H 250V
 (Cooper Bussmann part number S505H-1.25-R
 at www.cooperbussmann.com)



Caution Fuses located on bottom of device underneath door. Use Phillips #1 screwdriver for removal. Ensure all hazardous voltages are disconnected from the device prior to removal of door.



Fuse When this fuse symbol is marked on a device, take proper precautions.

Maximum common-mode voltage.....300 V DC or AC_{rms}

DC

Table 2. DC Voltage Accuracy

Range	Input Impedance	1-Year Accuracy ² ± (% of Reading + % of Range)	Temperature Coefficient ² ± (% of Reading + % of Range)/°C
100 mV*	>10 GΩ, 10 MΩ	0.015 + 0.005	0.001 + 0.0005
1 V	>10 GΩ, 10 MΩ	0.015 + 0.005	0.001 + 0.0005
10 V	>10 GΩ, 10 MΩ	0.015 + 0.005	0.001 + 0.0005
100 V	10 MΩ	0.035 + 0.005	0.005 + 0.0005
300 V	10 MΩ	0.035 + 0.005	0.005 + 0.0005

* With offset null.

Table 3. DC Current Accuracy

Range	Burden Voltage	1-Year Accuracy ² ± (% of Reading + % of Range)	Temperature Coefficient ² ± (% of Reading + % of Range)/°C
10 mA	<0.03 V	0.070 + 0.020	0.0035 + 0.0010
100 mA	<0.3 V	0.070 + 0.003	0.0020 + 0.0010
1 A	<0.03 V	0.130 + 0.025	0.0065 + 0.0010
10 A ⁹	<0.3 V	0.130 + 0.004	0.0045 + 0.0010

Table 4. DC Resistance Accuracy (2-Wire)^{*}, 1 V Open Circuit Voltage

Range	Short-Circuit Current	1-Year Accuracy ² ± (% of Reading + % of Range)	Temperature Coefficient ² ± (% of Reading + % of Range)/°C
100 Ω	170 μA	0.018 + 0.050	0.0010 + 0.0005
1 kΩ	170 μA	0.018 + 0.005	0.0010 + 0.0005
10 kΩ	70 μA	0.018 + 0.005	0.0010 + 0.0005
100 kΩ	10 μA	0.018 + 0.005	0.0010 + 0.0005
1 MΩ	1.1 μA	0.035 + 0.005	0.0040 + 0.0005
10 MΩ	1.1 μA	0.085 + 0.005	0.0100 + 0.0005
100 MΩ	1.1 μA	1.3 + 0.005	0.1000 + 0.0005

^{*} Perform offset nulling.



Caution The input terminals of the DMM are not protected for electromagnetic interference. As a result, the DMM may experience reduced measurement accuracy or other temporary performance degradation when connected to unshielded test leads in an environment with radiated or conducted radio frequency electromagnetic interference.

DC continuity accuracy range¹⁰.....100 Ω

DC diode test range.....2 V

⁹ 30 seconds on, 30 seconds off. Add 300 ppm/A for currents >2.2 A. After measuring >5 A, wait two minutes to get full accuracy in the 1 A range.

¹⁰ DC continuity is only available in the VirtualBench application.

Effective Common-Mode Rejection.....>100 dB
 Ratio (CMRR), 1 k Ω
 resistance in LO lead

Normal-Mode Rejection Ratio.....>100 dB
 (NMRR), 50/60 Hz \pm 0.1%

Overrange.....105% of range except 300 V

AC

Table 5. AC Voltage Accuracy

Range (rms)	Peak Voltage	Frequency	1-Year Accuracy ² \pm (% of Reading + % of Range)	Temperature Coefficient ² \pm (% of Reading + % of Range)/ $^{\circ}$ C
100 mV, 1 V, 10 V, 100 V, 265 V	\pm 210 mV, \pm 2.1 V, \pm 21 V, \pm 210 V, \pm 400 V	20 Hz to 45 Hz	0.91 + 0.10	0.01 + 0.005
		45 Hz to 65 Hz	0.30 + 0.05	0.01 + 0.005
		65 Hz to 1 kHz	0.21 + 0.05	0.01 + 0.005
		1 kHz to 5 kHz	0.12 + 0.05	0.01 + 0.005
		5 kHz to 20 kHz	0.35 + 0.05	0.01 + 0.005

Table 6. AC Current Accuracy

Range (rms)	Peak Current	Burden Voltage (rms)	Frequency	1-Year Accuracy ² \pm (% of Reading + % of Range)	Temperature Coefficient ² \pm (% of Reading + % of Range)/ $^{\circ}$ C
5 mA	\pm 10.5 mA	<0.02 V	20 Hz to 1 kHz	0.20 + 0.01	0.01 + 0.005
			1 kHz to 5 kHz	0.60 + 0.01	

Table 6. AC Current Accuracy (Continued)

Range (rms)	Peak Current	Burden Voltage (rms)	Frequency	1-Year Accuracy ² ± (% of Reading + % of Range)	Temperature Coefficient ² ± (% of Reading + % of Range)/°C
50 mA	±105 mA	<0.2 V	20 Hz to 1 kHz	0.20 + 0.01	0.01 + 0.005
			1 kHz to 5 kHz	0.50 + 0.01	
500 mA	±1.05 A	<0.02 V	20 Hz to 1 kHz	0.15 + 0.01	0.01 + 0.005
			1 kHz to 5 kHz	0.50 + 0.01	
5 A	±10.5 A	<0.2 V	20 Hz to 1 kHz	0.25 + 0.03	0.01 + 0.005
			1 kHz to 5 kHz	0.60 + 0.03	



Caution The input terminals of the DMM are not protected for electromagnetic interference. As a result, the DMM may experience reduced measurement accuracy or other temporary performance degradation when connected to unshielded test leads in an environment with radiated or conducted radio frequency electromagnetic interference.

Input impedance.....10 MΩ || 200 pF

CMRR, 1 kΩ resistance in LO lead.....>70 dB (DC to 60 Hz)

DC Power Supply

Outputs.....0 to +6 V/0 to 1 A,
 0 to +25 V/0 to 500 mA (isolated),
 0 to -25 V/0 to 500 mA (isolated)



Note The +25 V and -25 V channels are bank isolated from ground but not from each other.

Table 7. DC Accuracy/Resolution

Output	Type	+6 V	+25 V	-25 V
DC output ²	Voltage	0 to +6 V	0 to +25 V	0 to -25 V
	Current ¹¹	1 A	500 mA	500 mA
Programming accuracy ^{2,12} ± (% of reading + offset)	Voltage	0.1% + 5 mV	0.1% + 20 mV	0.1% + 20 mV
	Current	0.2% + 10 mA	0.15% + 4 mA	0.15% + 4 mA
Readback accuracy ^{2,12} ± (% of reading + offset)	Voltage	0.1% + 5 mV	0.1% + 20 mV	0.1% + 20 mV
	Current	0.2% + 10 mA	0.15% + 4 mA	0.15% + 4 mA
Programming resolution	Voltage	1.7 mV	6.5 mV	6.5 mV
	Current	0.30 mA	0.15 mA	0.15 mA
Readback resolution	Voltage	0.41 mV	1.7 mV	1.7 mV
	Current	70 µA	35 µA	35 µA
Load regulation ± (% of reading + offset) ¹³	Voltage	0.01% + 25 mV	0.03% + 5 mV	0.03% + 5 mV

Overvoltage protection.....30 V (25 V channels) and 10 V (6 V channel)

Reverse voltage protection.....400 mV

External Trigger (TRIG)

Direction control.....Input or output, software-selectable

Logic level.....5 V compatible LVTTTL input,
3.3 V LVTTTL output

Drive strength.....4 mA

Input voltage.....0 to 5 V



Note The external trigger line is designed to withstand accidental overvoltage from signals on the VB-8012 or similar devices. It is not recommended for use with signals likely to exceed 0 to 5 V in normal operation.

¹¹ Minimum programmable current limit is 1% of range.

¹² Programming and readback accuracy specified at no load.

¹³ Change in output voltage for any load within range.

Connectivity

Wired Interface

USB specification.....USB 2.0 Hi-Speed

Wireless Interface

Table 8. Network Protocols and Ports Used

Port	Protocol	Function
Port 80/TCP	HTTP	Device configuration (web, MAX)
Port 443/TCP	HTTP	Device configuration (web, MAX)
Port 3580/TCP	Service locator	Device configuration (web, MAX)
Port 9090/TCP	Configuration only	VirtualBench instrument protocol
Port 5353/UDP	Multicast DNS	Device discovery

Network IP configuration, date/time.....IPv4, DHCP Client/Server, Static IP

Radio mode.....IEEE 802.11 b,g,n

Wireless modes.....AP mode (default), client mode

Frequency band.....2.4 GHz ISM

Channel width.....20 MHz

Channels.....USA 1-11, International 1-13
(12 and 13 client mode only)

TX power.....+10 dBm max (10 mW)

Security.....Open, WPA, WPA2, WPA2-Enterprise

Enterprise security EAP types.....EAP-TLS, EAP-TTLS/MS-CHAPv2,
PEAPv0/MS-CHAPv2

Antenna.....External RP-SMA omnidirectional dipole

Power Requirements



Caution The protection provided by the VirtualBench hardware can be impaired if it is used in a manner not described in the *NI VB-8012 Safety, Environmental, and Regulatory Information* document.

Voltage input range.....	100 to 240 VAC, 50/60 Hz
Power consumption.....	100 W maximum
Power input connector.....	IEC C13 power connector
Power disconnect.....	The AC power cable provides main power disconnect. Do not position the equipment so that it is difficult to disconnect the power cable. Depressing the front panel power button does not inhibit the internal power supply.

Calibration

Calibration cycle (digital multimeter,..... mixed signal oscilloscope, function generator, DC power supply).....	1 year
Specified temperature.....	$T_{cal} \pm 5^{\circ}\text{C}$
Warmup time.....	30 minutes

Physical Characteristics

Dimensions

Enclosure.....	25.40 cm × 19.05 cm × 7.39 cm (10.00 in. × 7.50 in. × 2.91 in.)
Enclosure with connectors and..... antenna	25.40 cm × 23.37 cm × 13.97 cm (10.00 in. × 9.20 in. × 5.50 in.)



Note Use the VirtualBench instrument in a horizontal orientation. Allow at least 10.16 cm (4.0 in.) of clearance in front and behind the VirtualBench instrument for USB, power, and common connector cabling.

Weight.....	2.05 kg (4 lb 8.3 oz)
-------------	-----------------------

Connectivity

Mixed signal oscilloscope.....	BNC
Logic analyzer.....	2x20 shrouded IDC header
External trigger.....	BNC
Function generator.....	BNC
Digital I/O	
Type.....	Pluggable screw terminal, 3.5 mm (14 position)
Screw terminal wiring.....	0.1 to 2.0 mm ² (30 to 14 AWG)
Torque.....	0.25 N · m (2.2 lb · in.)
Digital multimeter.....	4 mm banana jacks
DC power supply	
Type.....	Pluggable screw terminal, 3.81 mm (6 position)
Screw terminal wiring.....	0.1 to 2.0 mm ² (30 to 14 AWG)
Torque.....	0.25 N · m (2.2 lb · in.)

If you need to clean the device, wipe it with a dry towel.

Safety Voltages

Connect only voltages that are within these limits.

DMM Isolation Voltages



Hazardous Voltage This icon denotes a warning advising you to take precautions to avoid electrical shock.

Channel-to-earth ground

Continuous.....	300 V, Measurement Category II
Withstand.....	3,000 V _{rms} , verified by a 5 s dielectric withstand test

Measurement Category II is for measurements performed on circuits directly connected to the electrical distribution system. This category refers to local-level electrical distribution, such as that provided by a standard wall outlet, for example, 115 V for U.S. or 230 V for Europe.



Caution Do not connect the VirtualBench hardware to signals or use for measurements within Measurement Categories III or IV.

DC Power Supply Isolation Voltages

+25 V and -25 V-to-earth ground.....60 VDC, Measurement Category I continuous



Note Measurement Categories CAT I and CAT O (Other) are equivalent. The input circuits are not intended for direct connection to the MAINS building installations of Categories CAT II, CAT III, or CAT IV.

RF Safety

This equipment complies with FCC radiation exposure limits set for uncontrolled equipment and meets the FCC radio frequency (RF) Exposure Guidelines in Supplement C to OET65. This product generates and radiates radio frequency energy. To comply with the radio frequency radiation exposure guidelines in an uncontrolled environment, this equipment should be installed and operated with at least 20 cm and more between the radiator and the person's body (excluding extremities: hands, wrists, feet, and legs).

This equipment complies with the European Council Recommendation (1995/519/EC) on the limitation of exposure of the general public to electromagnetic fields.

Environmental

Operating temperature.....0 to 40 °C

Storage temperature.....-20 to 70 °C

Operating humidity.....10 to 90% RH, noncondensing
DMM full accuracy at 10 to 80%

Storage humidity5 to 95% RH, noncondensing

Cooling.....Forced air circulation (positive pressurization) through a fan. Fan speed automatically adjusts according to operating conditions. Intake and exhaust locations are on rear of device. Ensure that the intake and exhaust locations are not obstructed.

Pollution Degree.....2

Maximum altitude.....2,000 m

Indoor use only.

Shock and Vibration

Operational shock.....	30 g peak, half-sine, 11 ms pulse (Tested in accordance with IEC 60068-2-27. Test profile developed in accordance with MIL-PRF-28800F.)
Random vibration	
Operating.....	5 to 500 Hz, 0.3 g _{rms}
Nonoperating.....	5 to 500 Hz, 2.4 g _{rms} (Tested in accordance with IEC-60068-2-64. Nonoperating test profile exceeds the requirements of MIL-PRF-28800F, Class 3.)

Safety

This product meets the requirements of the following standards of safety for electrical equipment for measurement, control, and laboratory use:

- IEC 61010-1, EN 61010-1
- UL 61010-1, CSA 61010-1



Note For UL and other safety certifications, refer to the product label or the [Online Product Certification](#) section.

Electromagnetic Compatibility

This product meets the requirements of the following EMC standards for radio equipment and for telecommunication terminal equipment:

- EN 61000-6-4 (IEC 61000-6-4): Emissions
- EN 61000-6-2 (IEC 61000-6-2): Immunity
- EN 55022 (CISPR 22): Class A emissions
- EN 55024 (CISPR 24): Immunity
- AS/NZS CISPR 22: Class A emissions
- FCC 47 CFR Part 15B: Class A emissions
- ICES-001: Class A emissions



Note For EMC declarations, certifications, and additional information, refer to the [Online Product Certification](#) section.

CE Compliance

This product meets the essential requirements of applicable European Directives, as follows:

- 2006/95/EC; Low-Voltage Directive (safety)
- 2004/108/EC; Electromagnetic Compatibility Directive (EMC)
- 1995/5/EC; Radio and Telecommunications Terminal Equipment (R&TTE) Directive

Online Product Certification

To obtain product certifications and the DoC for this product, visit ni.com/certification, search by model number or product line, and click the appropriate link in the Certification column.

Environmental Management

NI is committed to designing and manufacturing products in an environmentally responsible manner. NI recognizes that eliminating certain hazardous substances from our products is beneficial to the environment and to NI customers.

For additional environmental information, refer to the *Minimize Our Environmental Impact* web page at ni.com/environment. This page contains the environmental regulations and directives with which NI complies, as well as other environmental information not included in this document.

Waste Electrical and Electronic Equipment (WEEE)



EU Customers At the end of the product life cycle, all products must be sent to a WEEE recycling center. For more information about WEEE recycling centers, National Instruments WEEE initiatives, and compliance with WEEE Directive 2002/96/EC on Waste Electrical and Electronic Equipment, visit ni.com/environment/weee.

电子信息产品污染控制管理办法（中国 RoHS）



中国客户 National Instruments 符合中国电子信息产品中限制使用某些有害物质指令 (RoHS)。关于 National Instruments 中国 RoHS 合规性信息，请登录 ni.com/environment/rohs_china。(For information about China RoHS compliance, go to ni.com/environment/rohs_china.)

Refer to the *NI Trademarks and Logo Guidelines* at ni.com/trademarks for information on National Instruments trademarks. Other product and company names mentioned herein are trademarks or trade names of their respective companies. For patents covering National Instruments products/technology, refer to the appropriate location: **Help»Patents** in your software, the `patents.txt` file on your media, or the *National Instruments Patent Notice* at ni.com/patents. You can find information about end-user license agreements (EULAs) and third-party legal notices in the readme file for your NI product. Refer to the *Export Compliance Information* at ni.com/legal/export-compliance for the National Instruments global trade compliance policy and how to obtain relevant HTS codes, ECCNs, and other import/export data. NI MAKES NO EXPRESS OR IMPLIED WARRANTIES AS TO THE ACCURACY OF THE INFORMATION CONTAINED HEREIN AND SHALL NOT BE LIABLE FOR ANY ERRORS. U.S. Government Customers: The data contained in this manual was developed at private expense and is subject to the applicable limited rights and restricted data rights as set forth in FAR 52.227-14s, DFAR 252.227-7014, and DFAR 252.227-7015.

© 2014 National Instruments. All rights reserved.

371527A-01 Mar14